

Business Name: Dental Express Inc
Name: Wael Shahshouh
Address: 904 S. Roselle Rd. Unit 242
City: Schaumburg
State: IL
Zip: 60193
Phone: 8479244000
Fax:
Email: wael@dentalexpress.co

Comments/Notes: I replaced the handpiece bundle assembly, I replaced water solenoid. Checked voltage coming from board to solenoid is only 8.7 VDC, while it needs more to open the solenoid. Please check and call me for pricing to let my customer know before repair.

Equipment: Small Equipment Repair

Item #	Quantity	Make	Model	S/N	Problem
1	1	Little Beaver	2.0	A06357/2015	Not enough power to open solenoid

Tracking Code: 1Z65VV921207140442