

Business Name: Kelly S. Janecek DDS
Name: Kelly S Janecek
Address: 1498 South St. Francis Drive
City: Santa Fe
State: NM
Zip: 87505
Phone: 5059866073
Fax:
Email: Janecekdds@Gmail.com
Comments/Notes:

Equipment: Small Equipment Repair

Item #	Quantity	Make	Model	S/N	Problem
1	1	Dentsply Cavitron	SPS	124-56407	return water solenoid. replaced solenoid , water still not coming thru cavitron tips.

Tracking Code: 1Z65VV921201904306