

Business Name: Dental Fix

Name: Mark Belmar

Address: 12340 Seal Beach Blvd, Suite B653

City: Seal Beach

State: CA

Zip: 90740

Phone: 7143307881

Fax:

Email: mark.belmar@dentalfixrx.com

Comments/Notes: This Cavitron was service at your office on 9/28/2022 (see invoice# 5904) After the unit was sent back to me, I didn't use the unit. Two weeks ago, my customer bought it from me. They claim the tip has low vibration. When they increased the power to the tip, it stopped vibrating. Something is definitely wrong. I suspect the rebuild of the handpiece bundle may be the issue.

Equipment: Small Equipment Repair

Item #	Quantity	Make	Model	S/N	Problem
1	1	Dentsply Cavitron	Gen 119	119-31543	Office reported low vibration at tip.

Tracking Code: 1Z65VV921227659471