

Business Name: Reach Dental Equipment Service

Name: PJ Passey

Address: 1321 E Highland Ave

City: Carthage

State: MO

Zip: 64836

Phone: 4173163082

Fax:

Email: pjpassey@reachforservice.com

Comments/Notes: Unit will go through cycle without error. during "purge" cycle, when drive air is supposed to go through handpiece, nothing happens. Right side went out last year now left side will not work.

Equipment: Small Equipment Repair

Item #	Quantity	Make	Model	S/N	Problem
1	1	w&H	Assistina twin	05004	No Drive Air

Tracking Code: 1Z65VV921211136903