

Business Name: Fisher Dental

Name: Kili Maggard

Address: 328 S 2nd St

City: Coos Bay

State: OR

Zip: 97420

Phone: 5412692100

Fax:

Email: office.fisherdental@gmail.com

Comments/Notes: Dentsply Cavitron Jet Plus, Gen-132. The unit is not getting enough water through to the handpiece. Water supply to the unit is sufficient. Also checked water supply to the water solenoid, it was sufficient. I am pretty certain the problem lies in the handpiece cable, but I am not experienced w/ Cavitron repairs.

Equipment: Small Equipment Repair

Item #	Quantity	Make	Model	S/N	Problem
1	1	Dentsply	Cavitron Jet Plus GEN-132		The unit is not getting enough water to the handpiece

Tracking Code: 1Z65VV921217556903