

Business Name: 81st Street Dental

Name: Kathryn Beller

Address: 7711 E 81ST ST

City: Tulsa

State: OK

Zip: 74133-4200

Phone: 9184612766

Fax:

Email: kathryn@kathrynbellerdds.com

Comments/Notes: When the unit is turned on you can hear a high pitched sound and the water is coming out of the wand. The unit is not responding to the foot pedal or changing the dials on the unit. It will not vibrate when the handpiece is in it .

Equipment: Small Equipment Repair

Item #	Quantity	Make	Model	S/N	Problem
1	1	Dentsply	Cavitron Jet SPS Gen120	120-17904	Unit not responsive to the foot pedal

Tracking Code: 1Z65VV921204643351