

Business Name: Doster Dental
Name: Corinne Doster
Address: 1001 Carter St , Ste B
City: Chattanooga
State: TN
Zip: 37402
Phone: 4232658839

Fax:
Email: dentalcare@drdoster.net

Comments/Notes: Device is not vibrating properly when in use. Water still comes out and lights still come on. Also the handpiece operates with variable power depending on orientation of the handpiece (vertical vs horizontal, etc.)

Equipment: Handpiece Repair

Item #	Quantity	Make	Model	S/N	Problem
1	1	Parkell	TurboSensor Ultrasonic Scaler	54214	Ultrasonic tips don't vibrate as powerfully as it should.

Tracking Code: 1Z65VV920334915250