

Business Name: Doster Dental
Name: Corinne Doster
Address: 1001 Carter St , Ste B
City: Chattanooga
State: TN
Zip: 37402
Phone: 4232658839

Fax:
Email: dentalcare@drdoster.net

Comments/Notes: Ultrasonic tips don't vibrate as powerfully as it should. Also the handpiece power seems to vary depending on orientation of the handpiece (how it is being held....vertically or horizontally, etc.)

Equipment: Handpiece Repair

Item #	Quantity	Make	Model	S/N	Problem
1	1	Parkell	TurboSensor Ultrasonic Scaler	54214	Ultrasonic tips don't vibrate as powerfully as it should.

Tracking Code: 1Z65VV920328947273