

Business Name: myDental

Name: Edgar Jimenez

Address: 13450 N Highway 183, Suite 238A

City: Austin

State: TX

Zip: 78750

Phone: 5129008330

Fax:

Email: edgar.jimenez@mydental.com

Comments/Notes: Cavitron machine has been tested with other lines and is working fine. It is just the one tube that is needing repair.

Equipment: Handpiece Repair

Item #	Quantity	Make	Model	S/N	Problem
1	1	Dentsply	Cavitron 300 Series Tubing	G310-01747	Water is not coming out of the tubing. The machine works fine, only tubing repairs will be necessary.

Tracking Code: 1Z65VV921214538303