Business Name: Fix It Biomedical Name: Adam Burkhart Address: 14317 Potranco Rd, Ste 205/109 City: San Antonio State: TX Zip: 78245 Phone: 2104281190 Fax: Email: fixitbiomedical@gmail.com Comments/Notes: Box, faceplate, cord and r

Comments/Notes: Box, faceplate, cord and motor has been included. Adam has spoken with Paul about the issue that occured after he installed a cord that Kavo wired backward. Unit is in constant error. Unit was sent back to Kavo, but they said that the unit is old and a new unit needs to be purchased. Paul mentioned the possibility of installing a new/refurbished circuit board in it. Please advise on price before work is performed. Thank you!

Equipment: Small Equipment Repair

Item #	Quantity	Make	Model	S/N	Problem
1	1	Kavo	ELECTROtorque plus	2011-00018673	see below
2	1	Kavo	KL703LED	12-1005247	

Tracking Code: 1Z65VV921207340806